

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 23 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130307
Name		JAMES I. JENSEN, JR.	ID# 3393 Date 03-10-10
A	Agency	YAKUTAT DPS	Phone # 784-3206
Instrument Location AT POLICE STATION			
Alco S/N		77286	Target Value .082 High Pressure 505
Alco Test Values		1st Alco .081	2nd Alco .080
Signature		(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of April, 2010.

(Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130307 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130307 ✓

MARCH 13, 2010 ✓

OPERATOR'S NAME:

JENSEN JAMES I

OPERATOR'S NUMBER: 3393

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

L

O.J.L. #: L

DEPT/AGENCY: VAK1

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .002 ✓

ALCO S/N: 77286 ✓

### BREATH ANALYSIS

.002 ADJUSTED FOR 29.46 in ✓  
ALCO TARGET .000 19:40  
BLANK TEST .000 19:41  
INTERNAL STANDARD VERIFIED 19:41  
ALCO TV 29.47 in .001 ✓ 19:41  
BLANK TEST .000 19:42  
SUBJECT SAMPLE .000 19:43  
BLANK TEST .000 19:44  
ALCO TV 29.47 in .000 ✓ 19:44  
BLANK TEST .000 19:45

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130307 ✓

MARCH 13, 2010 ✓

TIME 19:37 ✓

### DIAGNOSTIC CHECK

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 40c  
BREATH TUBE: 41c  
BAROMETER: 29.46 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~||